

## ICP - Mass Spectrometry

## NexION® 1000G ICP-MS



The following performance can be measured under a single set of optimized conditions for each mode.

### Detection Limits

Based on three times the standard deviation of the blank using three-second integration time and peak hopping at 1-point per mass.

Element	Detection Limits ng/L (ppt)
<sup>9</sup> Be	< 0.5
<sup>59</sup> Co	< 0.5
<sup>115</sup> In	< 0.25

### Sensitivity

Element	Sensitivity Mcps/mg/L
<sup>9</sup> Be	> 6
<sup>115</sup> In	> 100
<sup>238</sup> U	> 80

## Oxide and Doubly-charged Species

Measured under identical operating conditions used to achieve sensitivity and detection-limit specifications.

CeO <sup>+</sup> /Ce <sup>+</sup>	< 0.025
Ce <sup>++</sup> /Ce <sup>+</sup>	< 0.03

## Background Signal

Mass 220	≤ 2 cps
----------	---------

## Short-term Precision

Defined as the relative standard deviation (% RSD) for a 1-10 µg/L multi-element solution, automatically cycling between Standard, Reaction and Collision modes, using a 3-second integration time, without internal standardization.

< 3% RSD
----------

## Long-term Stability

Relative stability after a one-hour warm-up period. Defined as the relative standard deviation of the mean signal for a 1-10 µg/L multi-element solution, automatically cycling between Standard, Reaction and Collision modes, measured once every 10 minutes, without internal standardization.

< 4% RSD over four hours
--------------------------

## Mass Calibration Stability

Measured using a 1 µg/L multi-element solution containing <sup>7</sup>Li, <sup>24</sup>Mg, and <sup>115</sup>In. Defined in terms of the shift in spectral position corresponding to maximum spectral peak intensity for each element, obtained without the use of multiple-point, peak-searching algorithms.

< 0.05 amu over eight hours of continuous operation
---

## Quadrupole Peak Hop (Slew) Speed

Defined as the maximum rate at which the quadrupole can jump over 160 amu without affecting the precision of the analytical measurement.

1.6 M amu/sec
---------------

## Quadrupole Scan Speed

Defined as the maximum rate at which the quadrupole can be scanned while acquiring continuous spectral data at every mass from the minimum to the maximum mass of the instrument (1-285 amu).

5000 amu/sec
--------------

## Detector Linear Range

The SimulScan™ detection system operates from < 0.1 cps to > 10<sup>9</sup> cps. This provides over 10 orders of magnitude of linear dynamic range in a single continuous scan.

## Transient Data Acquisition Speed

> 3000 temporal data points/sec maximum
Up to 100,000 temporal data points/sec in nano or single cell detection mode

JCANO INGENIERIA DE MÉXICO:  
Priv. Brunel Mz. 1 Lt. 5 Cs 4-B Fracc. Quinta Versalles  
C.P. 55767 Tecámac, Edo. De México  
Correo: [mcano@jcanoingenieria.com](mailto:mcano@jcanoingenieria.com)  
Página Web: <https://jcanoingenieria.com>  
Tel: 55 3996 2586  
WhatsApp: 55 7129 9832  
Facebook: <https://www.facebook.com/JCanoIngenieria>

PerkinElmer, Inc.  
940 Winter Street  
Waltham, MA 02451 USA  
P: (800) 762-4000 or  
(+1) 203-925-4602  
[www.perkinelmer.com](http://www.perkinelmer.com)



For a complete listing of our global offices, visit [www.perkinelmer.com/ContactUs](http://www.perkinelmer.com/ContactUs)

Copyright © 2019, PerkinElmer, Inc. All rights reserved. PerkinElmer® is a registered trademark of PerkinElmer, Inc. All other trademarks are the property of their respective owners.

014726A\_01

PKI